

GUIDANCE FOR COMPLETING SFDA MDMA APPLICATION FORM

AUSTRALIAN JURISDICTION (AU)

Medical Device Sector
Registration & Licenses Department

Revision: 20/06/2017
V 1.0

Contents

Section 1. Manufacturer (AU)	2
Section 2. General info. (AU)	3
Section 3. Jurisdiction. (AU)	11
Section 4. Product Categories (AU)	12
Section 5. Product Verification (AU)	13
Section 6. Manufacturer's QS Status (AU)	14
Section 7. Other National Provisions (AU)	15

Section 1. Manufacturer (AU)

No	SFDA Question	Do Check Warning	Task
1.1	Manufacturer	Do	<ul style="list-style-type: none"> Select the name of the manufacturer from the drop list
1.2	Legal Manufacturer	Do	<ul style="list-style-type: none"> Select the name of the Legal manufacturer.
		Check	<ul style="list-style-type: none"> The name and address of the devices' manufacturer in this application must concur with sections: <ul style="list-style-type: none"> 2.1.10 Labeling 2.1.11 IFU 2.3 A/C Power Supply Statement – if applicable 2.4 Environmental Statements 2.5 The provided documents in this section 5.3 EC certificates 5.3 Recent Audit report 5.3 Other Certificates as required by the device class 5.4 Declaration of conformity 6.3 QMS Certificate 7.1 Regulatory Compliance Attestation
		Warning	<ul style="list-style-type: none"> A common error is to select the device manufacturing site address, rather than the manufacturer address. If the manufacturer has two addresses a postal address and a Site address, please provide attested letter from the manufacturer explaining that there are two addresses – insert the letter in 2.1.10
1.3	Medical Device Category	Do	<ul style="list-style-type: none"> Use SFDA Drop-down list of 17 Categories

Section 2. General info. (AU)

No	SFDA Question	Do Check Warning	Task								
2.1	Details of the medical devices applying for marketing authorization	Do	<ul style="list-style-type: none"> Insert the list of devices in the application; make sure NOT to list the MODELS separately as different devices. For IVD, List the kit as a single item in section 2.1 for labels & IFU, all the labels of the reagent in the kit must be provided. 								
		Check	<ul style="list-style-type: none"> If the application contains multiple products then check that the listed products are meeting the bundling criteria. Refer to SFDA bundling rules in MDS-G7: http://www.sfda.gov.sa/en/medicaldevices/regulations/DocLib/MDS-G7.pdf Cross check the list of the devices against: Labels (2.1.10) IFU (2.1.11) If the device has multiple models, for example male urinary catheters of different sizes, the applicant should include these in one line: <table border="1" data-bbox="683 1066 1539 1251"> <thead> <tr> <th>Product description</th> <th>Intended Purpose</th> <th>Trade/Brand Name</th> <th>Model Number</th> </tr> </thead> <tbody> <tr> <td rowspan="3">Male urinary Catheters</td> <td rowspan="3">Drain urine from the bladder</td> <td rowspan="3">XYZ</td> <td>123-40</td> </tr> <tr> <td>123-50</td> </tr> <tr> <td>123-60</td> </tr> </tbody> </table> 	Product description	Intended Purpose	Trade/Brand Name	Model Number	Male urinary Catheters	Drain urine from the bladder	XYZ	123-40
Product description	Intended Purpose	Trade/Brand Name	Model Number								
Male urinary Catheters	Drain urine from the bladder	XYZ	123-40								
			123-50								
			123-60								
2.1.1	Product Brief Description (This field will appear on the MDMA printout)	Do	<ul style="list-style-type: none"> Insert the product brief description. Note: The product description will be printed on the MDMA license issued by the SFDA. 								
		Check	<ul style="list-style-type: none"> The product description must be precise and informative (Maximum of 100 Characters including spaces). The description must be in English only, no commas (acceptable if it makes sense "part of the sentence" e.g.: infant, pediatric & adult ventilator), clear & accurate. Make sure there is no spelling errors. 								

		Warning	<ul style="list-style-type: none"> • “Catheter, Urinary” will be rejected whereas “Urinary Catheter” is acceptable. • Do NOT include Brand Names or Company Names UNLESS the brand name is descriptive (Describing the product).
2.1.2	Intended Purpose of the medical device type	Do	<ul style="list-style-type: none"> • Insert the intended purpose.
		Check	<ul style="list-style-type: none"> • Typically this is an extract from IFU
2.1.3	Product Trade/Brand Name.	Do	<ul style="list-style-type: none"> • Insert the product Trade/Brand name as it appears on the label. • Note: The product Trade/Brand Name will be printed on the issued MDMA License.
		Check	<ul style="list-style-type: none"> • Check it concurs with the product trade/brand name as it appears on the product actual label.
		Warning	<ul style="list-style-type: none"> • The combination of the Product Description and Trade/Brand Name must be unique for every device listed in the application
2.1.4	Model Name/Number	Do	<ul style="list-style-type: none"> • Insert the model name/number as it appears on the label
		Check	<ul style="list-style-type: none"> • If more than one model number listed in the specified section, these models should only differ in color, size, weight, dimensions or shape.
		Warning	<ul style="list-style-type: none"> • If the product has model/ref number, the brand name must not be repeated in section 2.1
2.1.5	Manufacturer’s Device Identifier Number	Do	<ul style="list-style-type: none"> • Insert the Manufacturers Device Identifier Number
		Check	<ul style="list-style-type: none"> • Typically this is the REF number, or Product catalogue number. Check it concurs with the product ID number as it appears on the label.

2.1.6	Format of medical device identifier number(s) that will appear on labelling for traceability purposes	DO	<ul style="list-style-type: none"> Insert the Format of medical device identifier number(s) that will appear on the labeling <u>for traceability purposes</u> 							
		Check	<ul style="list-style-type: none"> Typically this is the LOT number, or Serial number Provide a brief description of how the number is formatted e.g. LOT YYYY-MM-DD (Year-Month-Day) 							
2.1.7 2.1.8 2.1.9	Nomenclature code number GMDN UMDNS Other(e.g. FDA identification number, JMDN)	DO	<ul style="list-style-type: none"> Insert the nomenclature code number if available. 							
2.1.10	Provide the label(s) affixed to the device or its wrappers when it is supplied to the KSA	DO	<ul style="list-style-type: none"> Attach the device labels for ALL devices listed in section 2.1 <p>A/C Power Supply If the product is connected to an A/C power supply, the applicant must provide images of the power supply label, so it can be cross-checked with the A/C Power Supply signed declaration provided in section 2.3 (60 Hz supply at nominal values of either 230 or 400 volts)</p> <ul style="list-style-type: none"> For IVD Kits, all the individual reagent labels must be provided. All labels must be in compliance with TGA and national provisions. 							
		Check	<ul style="list-style-type: none"> Check that labels provided for <u>ALL</u> the devices listed in section 2.1, including each of the models numbers/REF/Part No. /etc. When the device has range (e.g. sizes) then a representative label is acceptable provided that clearly links one product-size to one product ID number. <p>Example (Acceptable):</p> <table border="1"> <thead> <tr> <th>Trade Name</th> <th>REF (Product ID Number)</th> <th>Size (Product variable)</th> </tr> </thead> <tbody> <tr> <td>Medical Device</td> <td>1234</td> <td>5x5cm</td> </tr> <tr> <td>Medical Device</td> <td>1236</td> <td>10x10cm</td> </tr> </tbody> </table>	Trade Name	REF (Product ID Number)	Size (Product variable)	Medical Device	1234	5x5cm	Medical Device
Trade Name	REF (Product ID Number)	Size (Product variable)								
Medical Device	1234	5x5cm								
Medical Device	1236	10x10cm								

The applicant has provided a clear link between each of the product ID numbers and the product sizes/dimensions

- The table must be from the legal Manufacturer and must be signed with name, job title & date.
- The labels must contain:
 - 1- Device Trade name (See 2.1.3)
 - 2- Device model number (See 2.1.4)
 - 3- Device ID Number (REF) (See 2.1.5)
 - 4- Legal Manufacturers Name & Address (See 1.1 & 1.2)
- Labels May contain:
 - 1- LOT or Serial Number
 - 2- Power Supply – if applicable
 - 3- Storage Temperature
 - 4- Expiry Date
 - 5- Date of manufacture
 - 6- Sterile & method – if applicable
 - 7- Single use – if applicable
 - 8- The term “Made in ...” With the country of origin
 - 9- IVD – if applicable
 - 10- IVD Self test – if applicable
 - 11- Rx only – if applicable
- **Note: if the device is for professional use only:**
It is acceptable if the label provided in English only
Reference: SFDA MDS-IR6 Article 9 (C)
- **Note: If the device is for home use / Self Test (IVD)**
The label provided for each product must be written in both English & Arabic languages.
Reference: SFDA MDS-IR6 Article 9 (C)

Warning

- Common error is wrong or missing label.
- Tables, example (NOT Acceptable)

Trade Name	REF (Product ID Number)
Medical Device	1234
	1236
	etc

The applicant has NOT provided link between the product ID numbers and the product sizes/dimensions.

2.1.11	Provide Instructions For Use "IFU"	DO	Attach the IFU for ALL devices listed in section 2.1 and it comply with TGA
		Check	<ul style="list-style-type: none"> IFU cover ALL the devices Trade/Brand Names listed in section 2.1 Check that legal Manufacturers name & address are printed on the IFU and it concurs with sections 1.1 & 1.2 Check it contains Electrical rating –if applicable-. Note: if the device is for professional use only It is acceptable if the label provided in English only <u>Reference: SFDA MDS-IR6 Article 9 (C)</u> Note: If the device is for home use / Self Test (IVD) The label provided for each product must be written in both English & Arabic languages. <u>Reference: SFDA MDS-IR6 Article 9 (C)</u> Check it contains any WARNING &/or Precautions to take. Device's models & IDs (must match with 2.1 – if mentioned). If IFU does not cover all models or IDs or Doesn't match with sections 2.1.4 & 2.1.5, a justification is required. Storage Temperature: Min & Max (Where required) Sterility Method (Where required) Single use: (Where required) Power requirement if applicable (60 Hz Supply at nominal values or either 230 or 400 Volts) NOTE: IF IT IS NOT RELEVANT TO HAVE AN IFU FOR THE PRODUCT, THEN THE APPLICANT MUST PROVIDE A JUSTIFICATION FROM THE MANUFACTURER & MUST BE SIGNED, JOB TITTLE & DATED.
		Warning	<ul style="list-style-type: none"> A common error is a wrong or missing information OR missing IFU with no justification

2.1.12	List of Accessory	DO	<ul style="list-style-type: none"> List the accessory for the device (If applicable)
		Check	<ul style="list-style-type: none"> Accessory Definition: Accessories are devices specifically intended by its legal Manufacturer to be used together with the medical device to achieve its intended purpose.
		Warning	<ul style="list-style-type: none"> If the Accessory can be used as a stand-alone medical device, the SFDA do NOT consider it as an Accessory. It must be listed as a Medical Device.
2.1.12.1 to 2.1.12.11		DO	<ul style="list-style-type: none"> Same requirements as sections 2.1.1 to 2.1.11
2.2	Jurisdiction(s) Where this medical device may be placed on the market.	DO	<ul style="list-style-type: none"> Make selections as appropriate. Minimumly, AUSTRALIA (AU) must be selected because this is a AU submission.
2.3	Statement of Power Supply requirements for KSA Market	DO	<ul style="list-style-type: none"> If the device/accessory is connected to an a/c power supply, complete the statement Template provided (It must be printed on the <u>Manufacturer's Letterhead</u>) The statement will confirm the medical device is: <ol style="list-style-type: none"> Designed to operate with a 60 Hertz supply at nominal values of either 230 or 400 Volts. Is fitted with the appropriate a/c power connector. Maintains the required electrical safety conditions. Continues to perform to specification. (The statement must be signed, job tittle & dated from the manufacturer) If it is NOT applicable select the (N/A) box move to the next section.

		Check	<ul style="list-style-type: none"> The statement is signed, job title & dated from the manufacturer on its letterhead
		Warning	<ul style="list-style-type: none"> Do NOT alter wording of the SFDA template
2.4	Statement for KSA Environmental Factors	DO	<ul style="list-style-type: none"> Complete the statement template provided printed on Manufacturers letterhead and make sure it is signed, job title & dated.
		Check	<ul style="list-style-type: none"> The statement is signed, job title & dated from the manufacturer on its letterhead
		Warning	<ul style="list-style-type: none"> Do NOT alter wording of the SFDA template
2.5	Statement to correctly stored, transported, installed, maintained & disposed of medical devices	DO	<ul style="list-style-type: none"> Provide a copy of the manufacturer's instructions to ensure that the medical devices intended to be placed on the KSA market will be correctly stored, transported, installed, maintained & disposed of, and that users can be trained in their proper use and maintenance. The applicant may also provide additional information that they believe is relevant to this request. The additional information must be from the Manufacturer (Letterhead) and must be signed, job title & dated, also, list devices and accessories Trade/Brand Name in this application (See 2.1) or the application number.

2.6	Manufacturers advertising & marketing material intended for use in KSA	DO	<ul style="list-style-type: none"> Provide a copy of the manufacturer's advertising and marketing material intended for use in the KSA, if NOT AVAILABLE provide an <u>explanation</u> and the <u>date when such material will become available</u>, if NOT REQUIRED provide a <u>justification</u>. Note: It is acceptable to provide marketing materials that includes more devices than the listed in 2.1, however, the sections of the marketing material that includes the listed products should be HIGHLIGHTED Note: A product catalogue (Soft copy or hard copy) is considered marketing material and is acceptable.
		Check	<ul style="list-style-type: none"> Marketing literature is provided for at least ALL the devices listed in section 2.1 It should contain the manufacturer name Note: It is NOT necessary to have the address of the Manufacturer on the Marketing literature. It must contains the document control reference number Note: if the device is for professional use only It is acceptable if the marketing material provided in English only <u>Reference:</u> SFDA MDS-IR6 Article 9 (F) Note: If the device is for home use / Self Test (IVD) The Marketing material provided must be written in both English & Arabic languages. <u>Reference:</u> SFDA MDS-IR6 Article 9 (F)
		Warning	<ul style="list-style-type: none"> A common error is to state there is no marketing literature when a product catalogue is available A common error is to provide a marketing material without a document control reference number A common error when the Marketing literature includes more devices than the listed in section 2.1 and the relevant devices sections is NOT Highlighted. Marketing literature is NOT acceptable if provided on online-link or website.

Section 3. Jurisdiction. (AU)

No	SFDA Question	Do Check Warning	Task
3.1	Desired Jurisdiction	Do	Ensure AU jurisdiction has been selected.



Section 4. Product Categories (AU)

No	SFDA Question	Do Check Warning	Task
4.1	Device Type	Do	Select one correct option: <ul style="list-style-type: none"> • Medical Device • AIMD • IVD
4.2	Device Classification	Do	Select one correct option: <ul style="list-style-type: none"> • Medical device: <ol style="list-style-type: none"> 1- Class I 2- Class IIa 3- Class IIb 4- Class III • IVD: <ol style="list-style-type: none"> 1- Class I 2- Class II 3- Class III 4- Class IV
		Check	The selected Device classification is correct for the devices listed in 2.1, and it must concurs the classification stated in section 5 approvals & Certificates
4.3	(ONLY FOR IVD) Indicate whether the IVD is one of the following	Do	<ul style="list-style-type: none"> • Select the correct answer whether the IVD is 'listable', 'registrable' or 'Exempt' from the Australian Register of Therapeutic Goods.

Section 5. Product Verification (AU)

No	SFDA Question	Do Check Warning	Task
5.1	Medical device listing number as it appears on TGA's Register of Therapeutic goods	Do	<ul style="list-style-type: none"> Type the medical device listing number exactly as it appears on the TGA's Register of Therapeutic goods as evidence of registration.
5.2	Indicate the correct section number/s from the Therapeutic Goods Act applied to the device to establish conformity	DO	<ul style="list-style-type: none"> Type the sections number/s From the Therapeutic goods Act applied to the device to establish conformity to Australian regulations.
5.3	Provide a copy of the current valid certificate issued by TGA	DO	<ul style="list-style-type: none"> Provide a copy of the current certificates and most recent reports issued by this TGA as required by the indicated conformity assessment procedures.
5.4	Provide copy of the manufacturer current Declaration of conformity to the therapeutics goods Act.	DO	<ul style="list-style-type: none"> Provide the current DoC issued by the manufacturer make sure it is: <ul style="list-style-type: none"> signed, job tittle & dated from the manufacturer on its letterhead

Section 6. Manufacturer's QS Status (AU)

No	SFDA Question	Do Check Warning	Task
6.1	Indicate whether the manufacturer of the medical device operates an established quality management system (QMS) that complies with the required conformity assessment procedures	Do	<ul style="list-style-type: none"> Select the correct answer.
6.2	If YES, indicate the QMS standard used	Do	<ul style="list-style-type: none"> Type the standard used for QMS. Type the organization name.
6.3	the current quality management approvals/certificates held by the manufacturer	Do	<ul style="list-style-type: none"> Provide the current Quality Management Approval/Certificate held by the manufacturer, which relates to the listed products in section 2.1. Make sure it is valid.
6.4	Description of the medical devices covered by the QMS	Do	<ul style="list-style-type: none"> Write correct description of the medical device/s based on the QMS provided.
6.5	Indicate the procedure(s) that are included within the Manufacturer's QMS.	Do	<ul style="list-style-type: none"> Select the correct procedure and make sure it concur the QMS Provided.
6.6	If the device is a Class I medical device or a General IVD, and a QMS is operated but not required by the Therapeutic Drugs Act, is the QMS regularly audited by an independent organization?	Do	<ul style="list-style-type: none"> If relative & the answer is 'Yes' then type the name of the organization and provide a copy of the most recent QMS certificate issued by the independent organization.

Section 7. Other National Provisions (AU)

No	SFDA Question	Do Check Warning	Task
7.1	Manufacturer Declaration	Do	<ul style="list-style-type: none"> Complete the attestation using the SFDA template provided It must be printed on manufacturers Letterhead Select AU for Australian based application The attestation letter must be signed, name, job tittle and dated
7.2	Provide the address of the location where the manufacturer holds technical information to support this attestation	Do	<ul style="list-style-type: none"> Provide the address of the location where the manufacturer holds the technical information to support this attestation.
		Check	The FULL Postal address must be provided (Building, number, road, city, postal code, state, country)
7.3	Authorized Representative (AR) Declaration	Do	<ul style="list-style-type: none"> Complete the attestation using the SFDA template provided It must be printed on AR / Local Manufacturers (LM) Letterhead Write the correct application number The attestation letter must be signed with name, job tittle and dated in English and it must match the MDNR information (Arabic & English is Acceptable)